



Certificate number : 39330864  
Project number : 711247  
Page 1 of 2

Applicant Impulse Pumps B.V.  
Schooltinkweg 12  
7021 MC Zelhem  
The Netherlands

Submitted A liquid flow meter.

Manufacturer : Impulse  
Type : Digi-flow 4"  
Serial number : SMPL 0711001001  
Q<sub>max</sub> : 200 m<sup>3</sup>/h

The meter is provided with a digital display.

Calibration method The deviation of the meter as a function of flow rate has been determined by direct comparison with the Dutch National Standard for liquid quantity measurements(reference meters).  
Tests have been carried out using water with a pressure up to  $3.2 \cdot 10^5$  Pa and a mean temperature of 17.8°C ( $\pm 0.5^\circ\text{C}$ ).

Date of calibration 17 December 2007.

Results The results of the calibration are presented on page 2 of 2.

Traceability The results of the calibration services of NMI VSL are traceable to primary and/or (inter)nationally accepted measurement standards.

Dordrecht, 21 December 2007  
NMI Van Swinden Laboratorium B.V.

F.M. Smits  
Section Liquid Flow & Volume



*This certificate is consistent with Calibration and Measurement Capabilities (CMCs) that are included in Appendix C of the Mutual Recognition Arrangement (MRA) drawn up by the International Committee for Weights and Measures (CIPM). Under the MRA, all participating institutes recognize the validity of each other's calibration and measurement certificates for the quantities, ranges and measurement uncertainties specified in Appendix C (for details see <http://kcdb.bipm.fr>).*





### Results

Flow-rate [m <sup>3</sup> /h]	Deviation [%]
39.2	+0.5
78.7	+0.2
150.5	-0.1



The flow meter was not adjusted.

$$\text{Deviation [\%]} = \frac{\text{Indicated flow-rate} - \text{Reference flow-rate}}{\text{Reference flow-rate}} * 100 \%$$

The uncertainty in the deviation is less then or equal to 0.20%.  
 The reported uncertainty of measurement is based on the standard uncertainty of measurement multiplied by a coverage factor  $k = 2$ , which for a normal distribution corresponds to a coverage probability of approximately 95%. The standard uncertainty of measurement has been determined in accordance with the Guide to the Expression of Uncertainty in Measurement (GUM).